

Genesis Planning

Memorial Hermann North East Hybrid OR Expansion

Specification Room by Room Detail

Department: Surgery

Atta ID	Qty F/I	Description Model Spec Remarks / Item Notes	Manufacturer Vendor	Funding Source		Item Status Arch Code Custom Code	Water			Vacuum (lbs)		(in) Width Depth Height	Electrical/Mechanical		
				Arch Sig.	Spat. Sig.		Cold Hot Treated	Drain Steam Gas	Dental Medical Vent	Mounting Max Weight Ship Weight	Volts Amp Hz		Plug Type BTU/hr Ded. Circ.	Data Cnct. Emrg. Pwr.	
Room: Clean Core Room#: Room Sign: Area/Phase: Room Qty: 1															
Comments:															
3389-004	1	Autotransfusion Unit, General	Haemonetics Corporation (02005-110-	Project	Draft (New)	No	No	No	M	19.00	110	Type B (NEMA 5-15)			
	O/O	Cell Saver 5+	EP)	Yes	2-Movable, Elect	No	No	No	106	23.00	5	N/A	No	No	
		Dimensions include machine with cart (IV pole down). IV pole in full raised position makes height 91.50".													
5316-036	1	Cabinet, Warming, Dual, Freestanding	Getinge USA - Infection Control ()	Project	Draft (New)	No	No	No	F	30.00	115	Type B (NEMA 5-15)			
	O/C	5624 (24" Deep, Pedestal Base)	Getinge USA - Infection Control ()	Yes	2-Movable, Elect	No	No	No	303	24.00	12	1778	No	No	
		This model includes 6' power cord kit to replace standard hard wire connection. Requires dedicated electrical supply with accessible disconnect switch near cabinet. Quote Received													
5859-007	1	Cart, Procedure, Resuscitation	Armstrong Medical Industries, Inc. (AR-	Project	Draft (New)	No	No	No	M	39.00	N/A	N/A			
	O/O	AR-6 / EB-1	6 / EB-1)	No	3-Movable, Non-	No	No	No	155	25.00	N/A	N/A	No	No	
		Dimensions reflect approximate size with the AR-6 cart configured with the EB-1 package. Approximate dimensions with shelf and IV pole in "up" position: 86"H x 25"D x 51.5"W													
6015-001	4	Cart, Supply, Chrome, 48 inch	InterMetro Industries Corporation	Project	Draft (New)	No	No	No	M	48.00	N/A	N/A			
	O/V	Super Erecta 48x18x69 [N356EC]	(N356EC)	No	3-Movable, Non-	No	No	No	72	18.00	N/A	N/A	No	No	
		Yes Unassigned													
3678-041	1	Defibrillator, Monitor, w/Pacing	Physio-Control Inc. ()	Project	Draft (New)	No	No	No	C	10.50	120	Type B (NEMA 5-15)			
	O/O	LIFEPAK 20e (w/Pacing)	Physio-Control Inc. ()	No	2-Movable, Elect	No	No	No	12	10.50	1	N/A	No	No	
		No Unassigned													
6364-046	1	Dispenser, Glove, Triple Box	Sunbelt Plastic ()	Project	Draft (New)	No	No	No	W	10.25	N/A	N/A			
	O/C	GDSL-T	MMS Distributer ()	Yes	1-Fixed	No	No	No	2	4.00	N/A	N/A	No	No	
		No Unassigned													
3711-021	1	Dispenser, Medication, Auxiliary	CareFusion - Pyxis ()	Lease	Draft (New)	No	No	No	F	31.00	120	Type B (NEMA 5-15)			
	O/V	MedStation 3500 Sgl Column (4 doors)	CareFusion - Pyxis ()	Yes	2-Movable, Elect	No	No	No	314	28.00	<1	21	No	No	
		Data connection is to Medstation main console													
		No Unassigned													
3711-028	1	Dispenser, Medication, Auxiliary	CareFusion - Pyxis ()	Lease	Draft (New)	No	No	No	F	23.00	120	Type B (NEMA 5-15)			
	O/V	Pyxis MedStation 4000 (7-Drwr)	CareFusion - Pyxis ()	Yes	2-Movable, Elect	No	No	No	133	27.75	<1	21	No	No	
		Final weight of unit varies per drawer configuration. Data connection to Medstation Main console required.													
		No Unassigned													

Genesis Planning

Memorial Hermann North East Hybrid OR Expansion

Specification Room by Room Detail

Department: Surgery

Atta ID	Qty F/I	Description Model Spec Remarks / Item Notes	Manufacturer Vendor	Funding Source		Item Status Arch Code Custom Code	Water Vacuum (lbs)			(in) Width Depth Height	Electrical/Mechanical		
				Arch Sig. Spat. Sig.			Cold Hot Treated	Drain Steam Gas	Dental Medical Vent		Mounting Max Weight Ship Weight	Volts Amp Hz	Plug Type BTU/hr Ded. Circ.
3708-106	1	Dispenser, Medication, Host (Main)	CareFusion - Pyxis ()	Lease		Draft (New)	No	No	No	F	23.00	120	Type B (NEMA 5-15)
	O/V	Pyxis MedStation 4000 (6-Drwr, 0 Cubie)	CareFusion - Pyxis ()	Yes		2-Movable, Elect	No	No	No	166	27.00	<1	409 Yes
		Final weight of unit varies based on drawer configuration and addition of supplies. Electrical requirements are for main only. Each component requires independent outlet, see cutsheet for details. Dedicated circuit breaker recommended for all equipment.		No		Unassigned	No	No	No	166	55.00	60	Yes Yes
6451-003	1	Dispenser, Medication, Lock Module	CareFusion - Pyxis ()	Lease		Draft (New)	No	No	No	S	2.50	N/A	N/A
	O/V	SMART Remote Manager	CareFusion - Pyxis ()	Yes		3-Movable, Non-Elect	No	No	No	4	5.25	N/A	N/A No
		Mounts on refrigerator door. Connects to Pyxis Main dispenser for power. Requires 4" clearance on non-hinge side of refrigerator.		No		Unassigned	No	No	No	5	7.00	N/A	No No
3713-007	1	Dispenser, Supply, Auxiliary	CareFusion - Pyxis (S30A8DO0DR)	Lease		Draft (New)	No	No	No	F	52.00	110	Type B (NEMA 5-15)
	O/V	SupplyStation 30 Double Column		Yes		2-Movable, Elect	No	No	No	580	28.00	3	522 No
				No		Unassigned	No	No	No	580	79.50	60	No Yes
6452-001	1	Dispenser, Supply, Catheter	CareFusion - Pyxis (S30MB4DOCR)	Lease		Draft (New)	No	No	No	F	52.00	110	Type B (NEMA 5-15)
	O/O	CathRack Double Main		Yes		2-Movable, Elect	No	No	No	920	28.00	3	522 Yes
		Max weight reflects dispenser empty.		No		Unassigned	No	No	No	920	79.50	60	No No
3723-024	1	Disposal, Sharps, Wall Mount	Stericycle ()	Lease		Draft (New)	No	No	No	W	14.50	N/A	N/A
	O/V	Bio Systems C-02RES-0203-OC	Stericycle ()	Yes		1-Fixed	No	No	No	12	5.75	N/A	N/A No
		Max weight reflects cabinet with container full.		No		Unassigned	No	No	No	6	15.50	N/A	No No
6140-013	1	Monitor, Physiologic, Transport	GE Healthcare - Monitoring Systems	Project		Draft (New)	No	No	No	C	11.50	120	Type B (NEMA 5-15)
	O/O	Transport PRO	(Transport PRO)	No		2-Movable, Elect	No	No	No	9	6.00	<1	N/A Yes
				No		Unassigned	No	No	No	15	13.00	60	No No
4177-042	3	Pump, Infusion, Single	Baxter Healthcare - Sigma Pumps ()	Project		Draft (New)	No	No	No	C	4.25	120	Type B (NEMA 5-15)
	O/V	SIGMA Spectrum (w/Wireless Battery Module)	Baxter Healthcare - Sigma Pumps ()	No		2-Movable, Elect	No	No	No	2	2.50	<1	N/A No
		Electrical requirements are for battery recharging.		No		Unassigned	No	No	No	4	6.50	60	No No
3374-031	1	Pump, Suction/Aspirator, General,	SSCOR, Inc. ()	Project		Draft (New)	No	No	No	F	17.00	110	Type B (NEMA 5-15)
	O/O	Portable	Armstrong Medical Industries, Inc. ()	No		2-Movable, Elect	No	No	No	11	5.25	N/A	N/A No
		S-SCORT DUET		No		Unassigned	No	No	No	15	9.00	60	No No
4235-031	1	Refrigerator, Undercounter	Summit Appliance ()	Project		Draft (New)	No	No	No	F	23.75	115	Type B (NEMA 5-15)
	O/O	AL750SHIM (ADA Compliant/Shim)	Summit Appliance ()	Yes		2-Movable, Elect	No	No	No	95	24.50	1.3	N/A No
		Depth include handle. Height includes hinge cap.		No		Unassigned	No	No	No	102	32.50	60	No No

Genesis Planning

Memorial Hermann North East Hybrid OR Expansion

Specification Room by Room Detail

Department: Surgery

Atta ID	Qty F/I	Description Model Spec Remarks / Item Notes	Manufacturer Vendor	Funding Source		Water			Vacuum		(lbs) Mounting Max Weight Ship Weight	(in) Width Depth Height	Electrical/Mechanical		
				Arch Sig. Spat. Sig.	Item Status Arch Code Custom Code	Cold Hot Treated	Drain Steam Gas	Dental Medical Vent	Volts Amp Hz	Plug Type BTU/hr Ded. Circ.			Data Cnct. Emrg. Pwr.		
4363-011	3	Stand, IV, Multi-Pump	Pedigo Products ()	Project	Draft (New)	No	No	No	M	23.00	120	Type B (NEMA 5-15)			
	O/O	iSTAND P-1080-6 (w/Handle & Foley Hook)	Pedigo Products ()	No	2-Movable, Elect	No	No	No	47	23.00	N/A	N/A	No		
		Height adjustable from 69 to 124 inches. 120V/60Hz 6-outlet strip optional		No	Unassigned	No	No	No	47	124.00	60	No	No		
4688-004	1	Waste Can, Open Top	Rubbermaid Commercial Products ()	Project	Draft (New)	No	No	No	F	10.50	N/A	N/A	No		
	O/O	2543 Fire Resistant Beige (28 qt.)	MMS Distributer ()	No	3-Movable, Non-Elect	No	No	No	4	14.50	N/A	N/A	No		
				No	Unassigned	No	No	No	5	15.50	N/A	No	No		
Room: Control Room Room#: Room Sign: Area/Phase: Room Qty: 1															
Comments:															
3626-019	1	Clock, Elapsed Time, Wall Mount	American Time & Signal Co. ()	Project	Draft (New)	No	No	No	W	12.50	110	Hardwire			
	O/C	Digital Elapsed Timer (w/ Code Blue)	SQA261RSAE)	Yes	1-Fixed	No	No	No	4	3.00	<1	N/A	No		
		Requires SiteSync IQ system controller for time synchronization and Timer Control Station for setting and controls (both not included). 9V rechargeable ni-cad battery provides battery backup.	American Time & Signal Co. ()	No	Unassigned	No	No	No	6	6.50	60	No	Yes		
AJN977K	1	Clock, Wall, Synchronized, Analog	American Time & Signal Co. ()	Project	Draft (New)	No	No	No			N/A	N/A	No		
	O/O	SQ54BHDD904BP	American Time & Signal Co. ()	Yes	0-0-0-Information Technology	No	No	No			N/A	N/A	No		
				No	Unassigned	No	No	No			N/A	No	No		
7148-005	1	Controller, Clock, Synchronized, Wireless	American Time & Signal Co. (SSQMSTR-05N0E)	Project	Draft (New)	No	No	No	C	17.00	120	Hardwire			
	O/C	SiteSync IQ GPS Wireless System Kit (5 Watt)	American Time & Signal Co. (SSQMSTR-05N0E)	Yes	1-Fixed	No	No	No	8	8.00	<1	N/A	Yes		
		Dimensions reflect system controller only. Can be desk or rack mounted. Includes GPS antenna, should be mounted to outside wall or mast on roof with clear view of sky.		No	Unassigned	No	No	No	8	5.00	60	No	No		
7148-003	1	Controller, Clock, Synchronized, Wireless	American Time & Signal Co. (ATSTCS)	Project	Draft (New)	No	No	No	W	4.50	120	Hardwire			
	O/C	Timer Control Station (ATSTCS)	American Time & Signal Co. ()	Yes	1-Fixed	No	No	No	1	4.50	<1	N/A	No		
		Can be mounted up to 30 feet away from the Digital Clock/Timer.		No	Unassigned	No	No	No	1	0.25	60	No	No		
4688-001	2	Waste Can, Open Top	Rubbermaid Commercial Products ()	Project	Draft (New)	No	No	No	F	11.25	N/A	N/A	No		
	O/O	2544 Fire Resistant Beige (40 qt)	MMS Distributer ()	No	3-Movable, Non-Elect	No	No	No	6	15.00	N/A	N/A	No		
				No	Unassigned	No	No	No	8	20.00	N/A	No	No		

Room: Equip./ Electronics Room Room#: Room Sign: Area/Phase: Target Occupancy Room Qty: 1

Comments:

Genesis Planning

Memorial Hermann North East Hybrid OR Expansion

Specification Room by Room Detail

Department: Surgery

Atta ID	Qty F/I	Description Model Spec Remarks / Item Notes	Manufacturer Vendor	Funding Source		Water			Vacuum		(lbs) Mounting Max Weight Ship Weight	(in) Width Depth Height	Electrical/Mechanical		
				Arch Sig. Spat. Sig.	Item Status Arch Code Custom Code	Cold Hot Treated	Drain Steam Gas	Dental Medical Vent	Volts Amp Hz	Plug Type BTU/hr Ded. Circ.			Data Cnct. Emrg. Pwr.		
C-197443	1	Power Supply, Uninterruptible (UPS),	Philips Healthcare - Imaging Systems	Project	Draft (New)	No	No	No	F	51.00	480		N/A		
	O/C	3-Phase	()	No	2-Movable, Elect	No	No	No	2008	38.00	N/A	N/A	No		
		25 kVA/20 kW	Philips Healthcare - Imaging Systems ()	No	Unassigned	No	No	No		59.80	60	No	No		
Included in the Philips Flexmove quote.															
Room: Operating Room, Endo Vasc Hybrid Room#: Room Sign: Area/Phase: Room Qty: 1															
Comments:															
C-223923	1	Analyzer, Lab, Blood Gas, Point-of-	Abbott Point of Care Inc ()	Project	Draft (New)	No	No	No			N/A		N/A		
	O/V	Care	Abbott Point of Care Inc ()	No	2-Movable, Elect	No	No	No			N/A	N/A	No		
		i-Stat 300		No	Unassigned	No	No	No			N/A	No	No		
3370-047	1	Anesthesia Machine, General	GE Healthcare - Anesthesia Delivery	Project	Draft (New)	No	No	No	M	29.00	115		Type B (NEMA 5-15)		
	O/V	Aisys Carestation	(1011-9000-000)	Yes	2-Movable, Elect	No	No	Yes	370	35.00	15	N/A	Yes		
			GE Healthcare - Anesthesia Delivery	No	Unassigned	No	Yes	No	450	59.25	60	No	No		
15 Amp circuit required. Actual amp draw for anesthesia unit only is unavailable (per GE). Unit has on-board electrical outlets for peripherals. Changed model to Aisys. Aestiva no longer available. Changed price from 62,000 to 78,000. kn. 4-29-14															
7820-010	4	Apron, Lead Free	Infab Corporation (VASM-GL)	Project	Draft (New)	No	No	No	S	N/A	N/A		N/A		
	O/O	Vest and Skirt Male VASM (Greenlite)	P&C X-Ray Services, Inc. (VASM-GL)	No	3-Movable, Non-Elect	No	No	No	8	N/A	N/A	N/A	No		
				No	Unassigned	No	No	No	8	34.00	N/A	No	No		
7820-009	4	Apron, Lead Free	Infab Corporation (VASF-GL)	Project	Draft (New)	No	No	No	S	N/A	N/A		N/A		
	O/O	Vest and Skirt Female VASF (Greenlite)	P&C X-Ray Services, Inc. (VASF-GL)	No	3-Movable, Non-Elect	No	No	No	8	N/A	N/A	N/A	No		
				No	Unassigned	No	No	No	8	34.00	N/A	No	No		
4991-002	1	Atherectomy, Rotational, Coronary	Cardiovascular Systems, Inc. (DB-2000)	Project	Draft (New)	No	No	No	M	29.00	120		Type B (NEMA 5-15)		
	O/O	Diamondback 360		No	2-Movable, Elect	No	No	No	130	29.00	1	N/A	No		
				No	Unassigned	No	Yes	No	223	70.00	60	No	No		
C-127293	1	Board, Checklist, Perioperative	WorldWide Medical Solutions ()	Project	Draft (New)	No	No	No			N/A		N/A		
	C/C	Perioperative Checklist	WorldWide Medical Solutions ()	No	2-Movable, Elect	No	No	No			N/A	N/A	No		
				No	Unassigned	No	No	No			N/A	No	No		
C-127295	1	Board, OR Counts	WorldWide Medical Solutions ()	Project	Draft (New)	No	No	No			N/A		N/A		
	C/C	Counts Board	WorldWide Medical Solutions ()	No	2-Movable, Elect	No	No	No			N/A	N/A	No		
				No	Unassigned	No	No	No			N/A	No	No		

Genesis Planning

Memorial Hermann North East Hybrid OR Expansion

Specification Room by Room Detail

Department: Surgery

Atta ID	Qty F/I	Description Model Spec Remarks / Item Notes	Manufacturer Vendor	Funding Source		Water			Vacuum		(lbs) Mounting Max Weight Ship Weight	(in) Width Depth Height	Electrical/Mechanical		
				Arch Sig. Spat. Sig.	Item Status Arch Code Custom Code	Cold Hot Treated	Drain Steam Gas	Dental Medical Vent	Volts Amp Hz	Plug Type BTU/hr Ded. Circ.			Data Cnct. Emrg. Pwr.		
6400-035	1	Boom, Anesthesia	Skytron ()	Project	Draft (New)	No	No	No	CE	196.50	120		Hardwire		
	O/V	ET2FVBM36/2AFC2	Skytron ()	Yes	1-Fixed	No	No	Yes	809	196.50	20		N/A	Yes	
		Utilities connectors, electrical specifications, and detailed fixture configuration available in Drawing Package provided by Manufacturer.		No	Unassigned	No	Yes	No	1000	*	60		Yes	Yes	
C-187328	1	Boom, Equipment, Dual Arm	Skytron ()	Project	Draft (New)	No	No	No	CE		0		N/A		
	O/V	ET2F/2AF	Skytron ()	Yes	2-Movable, Elect	No	No	No			N/A		N/A	Yes	
				No	Unassigned	No	No	No			N/A		No	No	
C-187334	2	Boom, Imaging Monitor	Skytron ()	Project	Draft (New)	No	No	No	CE		0		N/A		
	O/V	EC2E	Skytron ()	Yes	2-Movable, Elect	No	No	No			N/A		N/A	No	
				No	Unassigned	No	No	No			N/A		No	No	
C-187329	1	Boom, X-Ray Shield	Skytron ()	Project	Draft (New)	No	No	No	CE		0		N/A		
	O/V	LC2AF/NL	Skytron ()	Yes	2-Movable, Elect	No	No	No			N/A		N/A	No	
				No	Unassigned	No	No	No			N/A		No	No	
3451-001	2	Bucket, Kick	Pedigo Products ()	Project	Draft (New)	No	No	No	M	13.50	N/A		N/A		
	O/O	P-1020-SS	Pedigo Products ()	No	3-Movable, Non-Elect	No	No	No	8	13.50	N/A		N/A	No	
				No	Unassigned	No	No	No	8	14.00	N/A		No	No	
C-153741	2	C-View Integrated Display System	Carrot Medical ()	Project	Draft (New)	No	No	No			N/A		N/A		
	O/V	56 in	Philips Healthcare - Imaging Systems ()	Yes	2-Movable, Elect	No	No	No			N/A		N/A	No	
		Included with Philips X-Ray System.		No	Unassigned	No	No	No			N/A		No	No	
5594-081	5	Cabinet, OR Console, Supply	InnerSpace - Datel (514818DGAS)	Project	Draft (New)	No	No	No	F	49.50	N/A		N/A		
	O/C	514818DGAS Stainless Steel	MMS Distributer (514818DGAS)	Yes	1-Fixed	No	No	No	325	18.75	N/A		N/A	No	
				No	Unassigned	No	No	No	325	84.00	N/A		No	No	
5594-012	4	Cabinet, OR Console, Supply	Getinge USA - Infection Control ()	Project	Draft (New)	No	No	No	F	48.00	N/A		N/A		
	O/C	48" OR Console (w/sloping top & sub-base)	Getinge USA - Infection Control ()	Yes	1-Fixed	No	No	No	380	18.00	N/A		N/A	No	
		Height includes 8" sloping top and 12" sub-base (optional). Quote Received		No	Unassigned	No	No	No	470	80.00	N/A		No	No	
6815-027	1	Camera, Endoscopy, Control Unit	Stryker Endoscopy (1488-010-001)	Project	Draft (New)	No	No	No	C	12.50	115		Type B (NEMA 5-15)		
	O/O	1488 Camera Control Unit (Fiber Output)		No	2-Movable, Elect	No	No	No	13	15.25	<1		N/A	No	
		Equipment Boom		No	Unassigned	No	No	No	16	4.00	60		No	No	
6824-002	1	Cart, Anesthesia, Controlled Access	CareFusion - Pyxis ()	Lease	Draft (New)	No	No	No	M	47.50	120		Type B (NEMA 5-15)		
	O/O	Anesthesia System	CareFusion - Pyxis ()	Yes	2-Movable, Elect	No	No	No	400	28.00	3		355	Yes	
		Dimensions specified reflect unit fully configured with return bin and top accessory rack. Requires network connection if wireless option not selected. Consult Vendor for project-specific configuration and specification.		No	Unassigned	No	No	No	450	61.00	60		No	No	

Genesis Planning

Memorial Hermann North East Hybrid OR Expansion

Specification Room by Room Detail

Department: Surgery

Atta ID	Qty F/I	Description Model Spec Remarks / Item Notes	Manufacturer Vendor	Funding Source		Water			Vacuum (lbs)		(in) Width Depth Height	Electrical/Mechanical		
				Arch Sig. Spat. Sig.	Item Status Arch Code Custom Code	Cold Hot Treated	Drain Steam Gas	Dental Medical Vent	Mounting Max Weight Ship Weight	Volts Amp Hz		Plug Type BTU/hr Ded. Circ.	Data Cnct. Emrg. Pwr.	
3635-009	1 O/O	Compression Unit, Extremity Pump, Intermittent Venaflo 30A	DJO Global, Inc. () DJO Global, Inc. ()	Lease	Draft (New)	No	No	No	C	11.00	115	Type B (NEMA 5-15)		
				No	2-Movable, Elect	No	No	No	9	5.00	N/A	N/A	No	
				No	Unassigned	No	No	No	13	11.00	60	No	No	
3645-012	1 O/V	Computer Workstation, Cardiac Cath- Lab, Hemodynamic MacLab IT There is an existing system in the hospital (Witt System). This system is no longer available. Hospital has not made a decision on which system to use hospital wide , so asked for this system to be stand alone.	GE Healthcare - Cardiology () GE Healthcare - Cardiology ()	Project	Draft (New)	No	No	No	S	65.00	115	Type B (NEMA 5-15)		
				No	2-Movable, Elect	No	No	No	172	30.00	15	N/A	Yes	
				No	Unassigned	No	No	No	N/A	30.00	60	No	No	
3713-017	1 O/V	Dispenser, Supply, Auxiliary SupplyStation 30 Single Column	CareFusion - Pyxis () CareFusion - Pyxis ()	Lease	Draft (New)	No	No	No	F	31.00	110	Type B (NEMA 5-15)		
				Yes	2-Movable, Elect	No	No	No	300	28.00	3	490	No	
				No	Unassigned	No	No	No	300	79.50	60	No	Yes	
6452-001	1 O/O	Dispenser, Supply, Catheter CathRack Double Main Max weight reflects dispenser empty.	CareFusion - Pyxis () CareFusion - Pyxis ()	Lease	Draft (New)	No	No	No	F	52.00	110	Type B (NEMA 5-15)		
				Yes	2-Movable, Elect	No	No	No	920	28.00	3	522	Yes	
				No	Unassigned	No	No	No	920	79.50	60	No	No	
3712-014	1 O/V	Dispenser, Supply, Host StockStation (w/Wireless Scanner)	CareFusion - Pyxis () CareFusion - Pyxis ()	Lease	Draft (New)	No	No	No	F		0	N/A		
				Yes	2-Movable, Elect	No	No	No	N/A		N/A	N/A	No	
				No	Unassigned	No	No	No	N/A	*	N/A	No	Yes	
6457-020	1 O/V	Disposal, Sharps, Floor Cart BioSystems C-17/DWSH-17/LH-17	Stericycle () Stericycle ()	Lease	Draft (New)	No	No	No	M	14.50	N/A	N/A		
				No	3-Movable, Non- Elect	No	No	No	15	22.50	N/A	N/A	No	
				No	Unassigned	No	No	No	18	29.75	N/A	No	No	
7823-005	1 O/V	Docking Station, Analyzer, Blood Gas, POC i-STAT 1 Downloader Recharger Kit (DRC-300) Needs to be able to innerface. Contact John Sims with Telcor for quote. knd.	Abbott Point of Care Inc (04P73-04) Abbott Point of Care Inc (04P73-04)	Project	Draft (New)	No	No	No	C	4.25	120	Type B (NEMA 5-15)		
				Yes	2-Movable, Elect	No	No	No	2	9.75	1.1	N/A	Yes	
				No	Unassigned	No	No	No	3	5.00	60	No	No	
C-119886	1 O/O	Docking Station, iPod TBD	TBD () TBD ()	Project	Draft (New)	No	No	No			0	N/A		
				No	2-Movable, Elect	No	No	No			N/A	N/A	No	
				No	Unassigned	No	No	No			N/A	No	No	
6145-006	2 O/O	Doppler, Vascular Dopplex D900 w/ EZ8 Probe Requires 9 volt alkaline battery-6LR61, 6LF22 or equivalent.	ArjoHuntleigh () Huntleigh Healthcare, Inc ()	Project	Draft (New)	No	No	No	C	3.00	N/A	N/A		
				No	3-Movable, Non- Elect	No	No	No	1	1.25	N/A	N/A	No	
				No	Unassigned	No	No	No	2	5.50	N/A	No	No	

Genesis Planning

Memorial Hermann North East Hybrid OR Expansion

Specification Room by Room Detail

Department: Surgery

Atta ID	Qty F/I	Description Model Spec Remarks / Item Notes	Manufacturer Vendor	Funding Source Arch Sig. Spat. Sig.	Item Status Arch Code Custom Code	Water			Vacuum (lbs)		(in) Width Depth Height	Electrical/Mechanical		
						Cold Hot Treated	Drain Steam Gas	Dental Medical Vent	Mounting Max Weight Ship Weight	Volts Amp Hz		Plug Type BTU/hr Ded. Circ.	Data Cnct. Emrg. Pwr.	
4824-060	1 O/V	Electrosurgical Unit, Bipolar ForceTriad	Covidien - Valleylab Div () Covidien - Valleylab Div ()	Project	Draft (New)	No	No	No	C	18.00	120	Type B (NEMA 5-15)		
				Yes	2-Movable, Elect	No	No	No	30	20.00	7	N/A	No	
				Yes	Unassigned	No	No	No	30	10.00	60	No	No	
5618-005	1 O/O	Electrosurgical Unit, Monopolar Peak PlasmaBlade System (w/ Cart) Dimensions include generator with cart and are estimated. Actual cart dimensions unavailable from manufacturer at this time.	Medtronic Advanced Energy (PS100-100/PS100-300)	Project	Draft (New)	No	No	No	M	20.00	115	Type B (NEMA 5-15)		
				No	2-Movable, Elect	No	No	No	45	20.00	5.3	N/A	No	
				No	Unassigned	No	No	No	45	40.00	60	No	No	
3806-054	4 O/O	Flowmeter, Air 7702-1264-921	Ohio Medical () MMS Distributer ()	Project	Draft (New)	No	No	No	W	1.25	N/A		N/A	
				No	3-Movable, Non-Elect	No	No	No	1	3.50	N/A	N/A	No	
				No	Unassigned	No	Yes	No	1	5.50	N/A	No	No	
3803-078	7 O/O	Flowmeter, Oxygen 7702-1260-921	Ohio Medical () MMS Distributer ()	Project	Draft (New)	No	No	No	W	1.25	N/A		N/A	
				No	3-Movable, Non-Elect	No	No	No	1	3.50	N/A	N/A	No	
				No	Unassigned	No	Yes	No	1	5.50	N/A	No	No	
3836-009	3 O/O	Hamper, Linen P-120-L	Pedigo Products () Pedigo Products ()	Project	Draft (New)	No	No	No	M	19.50	N/A		N/A	
				No	3-Movable, Non-Elect	No	No	No	24	18.00	N/A	N/A	No	
				No	Unassigned	No	No	No	24	32.00	N/A	No	No	
7399-002	1 O/O	Headlight, LED ri-focus 6092 Electrical requirement is for battery charging.	Riester USA (6092)	Project	Draft (New)	No	No	No	C	11.00	120	Type B (NEMA 5-15)		
				No	3-Movable, Non-Elect	No	No	No	2	8.00	N/A	N/A	No	
				No	Unassigned	No	No	No	2	6.00	60	No	No	
3908-021	1 O/O	Injector, Contrast Media, Mobile Mark 7 Arterion Connects to power unit (15.21"D x 9.76"W x 4.64"H, 11 lbs.)	Medrad Inc (ART 700 PEDL) Medrad Inc (ART 700 PEDL)	Capital	Draft (New)	No	No	No	M	22.00	120		Hardwire	
				Yes	2-Movable, Elect	No	No	No	146	47.50	2	N/A	No	
				No	Unassigned	No	No	No	146	57.50	60	No	No	
3910-016	1 O/O	Insufflator, CO2 PneumoSure XL (45L) Equipment Boom	Stryker Endoscopy (0620-040-610) Stryker Endoscopy (0620-040-610)	Project	Draft (New)	No	No	No	C	12.50	120	Type B (NEMA 5-15)		
				No	2-Movable, Elect	No	No	No	20	18.75	3.15	N/A	No	
				No	Unassigned	No	No	No	25	6.00	60	No	No	
C-197608	1 O/V	Integration System, Surgical, Allowance TBD	Perkins () Perkins ()	Project	Draft (New)	No	No	No			N/A		N/A	
				No	2-Movable, Elect	No	No	No			N/A	N/A	No	
				No	Unassigned	No	No	No			N/A	No	No	
3952-035	1 O/O	Light Source, Xenon X8000 Equipment Boom	Stryker Endoscopy (0220-200-000) Stryker Endoscopy (0220-200-000)	Project	Draft (New)	No	No	No	C	12.50	120	Type B (NEMA 5-15)		
				No	2-Movable, Elect	No	No	No	16	17.00	3.8	N/A	No	
				No	Unassigned	No	No	No	21	4.75	60	No	No	

Genesis Planning

Memorial Hermann North East Hybrid OR Expansion

Specification Room by Room Detail

Department: Surgery

Atta ID	Qty F/I	Description Model Spec Remarks / Item Notes	Manufacturer Vendor	Funding Source Arch Sig. Spat. Sig.	Item Status Arch Code Custom Code	Water			Vacuum		(lbs) Mounting Max Weight Ship Weight	(in) Width Depth Height	Electrical/Mechanical		
						Cold Hot Treated	Drain Steam Gas	Dental Medical Vent	Volts Amp Hz	Plug Type BTU/hr Ded. Circ.			Data Cnct. Emrg. Pwr.		
C-187333	1	Light, Surgical, Dual, Ceiling	Skytron ()	Project	Draft (New)	No	No	No	CE			0		N/A	
	O/V	Aurora II LED AUR55TVC-200/1500	Skytron ()	Yes	2-Movable, Elect	No	No	No					N/A	No	No
				No	Unassigned	No	No	No					N/A	No	No
5046-091	1	Light, Surgical, Single, Ceiling	Skytron ()	Project	Draft (New)	No	No	No	CE	178.50	120		Hardwire		
	O/V	Aurora II LED AUR5	Skytron ()	Yes	1-Fixed	No	No	No	132	178.50	1.6	490	No	No	
		Wall mounted intensity control.		No	Unassigned	No	No	No	180	*	60	Yes	No	No	
4076-005	1	Monitor, Physiologic, Bedside	GE Healthcare - Monitoring Systems ()	Project	Draft (New)	No	No	No	C	16.00	120		Type B (NEMA 5-15)		
	O/V	Dinamap MPS Select	GE Healthcare - Monitoring Systems ()	Yes	2-Movable, Elect	No	No	No	40	12.50	N/A	N/A	No	No	
				No	Unassigned	No	No	No	N/A	15.00	60	No	No	No	
4072-004	1	Monitor, Physiologic, EEG,	Covidien RMS (Respiratory &	Project	Draft (New)	No	No	No	C	7.50	120		Type B (NEMA 5-15)		
	O/O	Intraoperative	Monitoring) ()	No	2-Movable, Elect	No	No	No	5	5.00	<1	N/A	No	No	
		BIS Vista Monitor	Covidien - Respiratory and Monitoring ()	No	Unassigned	No	No	No	8	8.00	60	No	Yes		
C-243866	2	Monitor, Video, 26 - 32 inch, Medical	Panasonic ()	Project	Draft (New)	No	No	No				N/A		N/A	
	O/O	Grade	Skytron ()	Yes	2-Movable, Elect	No	No	No				N/A	N/A	No	No
		26" LED Backlit IPS-PRO HD DISPLAY		No	Unassigned	No	No	No				N/A	No	No	No
		Part of Skytron quote.													
C-108241	1	Power Pack Rechargeable i-STAT	Abbott Point of Care Inc (06F23-55)	Project	Draft (New)	No	No	No				N/A		N/A	
	O/V	06F23-55	Abbott Point of Care Inc (06F23-55)	Yes	2-Movable, Elect	No	No	No				N/A	N/A	No	No
				No	Unassigned	No	No	No				N/A	No	No	No
C-153425	1	Pre-Installation Equipment, Lights and	Skytron ()	Project	Draft (New)	No	No	No				N/A		N/A	
	O/C	Booms	Skytron ()	Yes	2-Movable, Elect	No	No	No				N/A	N/A	No	No
		TBD		No	Unassigned	No	No	No				N/A	No	No	No
4187-048	1	Rack, Apron, Wall Mount	Infab Corporation (683411)	Project	Draft (New)	No	No	No	W	46.50	N/A		N/A		
	O/C	683411 (8 peg)	P&C X-Ray Services, Inc. (683411)	No	1-Fixed	No	No	No	18	7.25	N/A	N/A	No	No	
				No	Unassigned	No	No	No	18	9.00	N/A	No	No	No	
7910-001	1	Rack, Sponge Counting	Xodus Medical (50511)	Project	Draft (New)	No	No	No	C	12.00	N/A		N/A		
	O/O	50511 Double Sided (I.V. Pole)	Covidien - Kendall Products (31144184)	No	3-Movable, Non-Elect	No	No	No	1	6.00	N/A	N/A	No	No	
				No	Unassigned	No	No	No	2	12.00	N/A	No	No	No	
6618-042	1	Recorder, CD/DVD, Medical	Stryker Endoscopy (0240-060-921)	Project	Draft (New)	No	No	No	C	12.50	115		Type B (NEMA 5-15)		
	O/O	SDC3 Base w/DICOM and SDP1000 Printer Kit	Stryker Endoscopy (0240-060-921)	Yes	2-Movable, Elect	No	No	No	20	16.50	4	N/A	Yes		
		Equipment Boom		No	Unassigned	No	No	No	25	7.25	60	No	No	No	

Genesis Planning

Memorial Hermann North East Hybrid OR Expansion

Specification Room by Room Detail

Department: Surgery

Atta ID	Qty F/I	Description Model Spec Remarks / Item Notes	Manufacturer Vendor	Funding Source Arch Sig. Spat. Sig.	Item Status Arch Code Custom Code	Water			Vacuum (lbs)		(in) Width Depth Height	Electrical/Mechanical		
						Cold Hot Treated	Drain Steam Gas	Dental Medical Vent	Mounting Max Weight Ship Weight	Volts Amp Hz		Plug Type BTU/hr Ded. Circ.	Data Cnct. Emrg. Pwr.	
4248-085	8	Regulator, Suction,	Ohio Medical ()	Project	Draft (New)	No	No	No	W	3.00	N/A		N/A	
	O/O	Intermittent/Continuous 8702-1251-908 Push-To-Set Analog	MMS Distributer ()	No	3-Movable, Non-Elect	No	No	Yes	2	5.00	N/A	N/A	No	
				No	Unassigned	No	No	No	2	6.75	N/A	No	No	
4254-006	1	Roller, Patient Transfer	Alimed, Inc. (920890)	Project	Draft (New)	No	No	No	W	14.00	N/A		N/A	
	O/O	920890 (Medium Roller/Rack Kit) Dimensions reflect transfer device hanging on device rack.	MMS Distributer (920890)	No	3-Movable, Non-Elect	No	No	No	10	1.75	N/A	N/A	No	
				No	Unassigned	No	No	No	10	30.00	N/A	No	No	
5686-001	1	Services, Installation	To Be Determined ()	Project	Draft (New)	No	No	No	N/A	N/A	N/A		N/A	
	O/V	TBD	To Be Determined (Stryker)	No	0-0-0-Information	No	No	No	N/A	N/A	N/A	N/A	No	
				No	Technology Unassigned	No	No	No	N/A	N/A	N/A	No	No	
4311-009	1	Shield, Lead, Mobile	Wolf X-Ray Corporation (56604)	Project	Draft (New)	No	No	No	M	30.00	N/A		N/A	
	O/O	56604 Clear Pb	P&C X-Ray Services, Inc. (56604)	No	3-Movable, Non-Elect	No	No	No	100	24.00	N/A	N/A	No	
				No	Unassigned	No	No	No	113	75.00	N/A	No	No	
5796-002	1	Stand, Basin, Double	Pedigo Products ()	Project	Draft (New)	No	No	No	M	28.00	N/A		N/A	
	O/O	P-1079-SS	Pedigo Products ()	No	3-Movable, Non-Elect	No	No	No	30	14.00	N/A	N/A	No	
				No	Unassigned	No	No	No	30	34.00	N/A	No	No	
5795-004	1	Stand, Basin, Single	Pedigo Products ()	Project	Draft (New)	No	No	No	M	14.00	N/A		N/A	
	O/O	P-1078-SS	Pedigo Products ()	No	3-Movable, Non-Elect	No	No	No	15	14.00	N/A	N/A	No	
				No	Unassigned	No	No	No	15	34.00	N/A	No	No	
4411-017	1	Stool, Anesthetist	Pedigo Products (P-6000/P-6000PR)	Project	Draft (New)	No	No	No	M	21.00	N/A		N/A	
	O/O	P-6000 w/Procedure Rest		No	3-Movable, Non-Elect	No	No	No	110	21.00	N/A	N/A	No	
				No	Unassigned	No	No	No	110	42.00	N/A	No	No	
4414-016	1	Stool, Exam, Cushion-Seat	Pedigo Products ()	Project	Draft (New)	No	No	No	M	23.50	N/A		N/A	
	O/O	P-526-GS	Pedigo Products ()	No	3-Movable, Non-Elect	No	No	No	16	23.50	N/A	N/A	No	
				No	Unassigned	No	No	No	16	23.75	N/A	No	No	
5002-001	4	Stool, Step, Stackable	Pedigo Products ()	Project	Draft (New)	No	No	No	F	21.00	N/A		N/A	
	O/O	P-1015	Pedigo Products ()	No	3-Movable, Non-Elect	No	No	No	15	16.00	N/A	N/A	No	
		Dimensions reflect width and depth with handles. Width and depth without handles: 19"W x 14"D		No	Unassigned	No	No	No	15	6.00	N/A	No	No	

Genesis Planning

Memorial Hermann North East Hybrid OR Expansion

Specification Room by Room Detail

Department: Surgery

Atta ID	Qty F/I	Description Model Spec Remarks / Item Notes	Manufacturer Vendor	Funding Source		Water			Vacuum		(lbs) Mounting Max Weight Ship Weight	(in) Width Depth Height	Electrical/Mechanical		
				Arch Sig. Spat. Sig.	Item Status Arch Code Custom Code	Cold Hot Treated	Drain Steam Gas	Dental Medical Vent	Volts Amp Hz	Plug Type BTU/hr Ded. Circ.			Data Cnct. Emrg. Pwr.		
4424-004	1 O/O	Stool, Surgeon P-7000 Hydraulic Surgeon Chair	Pedigo Products (P-7000) MMS Distributer (P-7000)	Project	Draft (New)	No	No	No	M	21.00	N/A		N/A		
				No	3-Movable, Non-Elect	No	No	No	120	21.00	N/A	N/A	No		
				No	Unassigned	No	No	No	120	45.00	N/A	No	No		
5914-007	1 O/O	Table, Instrument, 60 inch SG-94-SS (60x24)	Pedigo Products () Pedigo Products ()	Project	Draft (New)	No	No	No	M	60.00	N/A		N/A		
				No	3-Movable, Non-Elect	No	No	No	88	24.00	N/A	N/A	No		
				Yes	Unassigned	No	No	No	88	34.00	N/A	No	No		
5990-005	2 O/O	Table, Utility, General Purpose SG-81-SS	Pedigo Products () Pedigo Products ()	Project	Draft (New)	No	No	No	M	20.00	N/A		N/A		
				No	3-Movable, Non-Elect	No	No	No	37	16.00	N/A	N/A	No		
				No	Unassigned	No	No	No	37	34.00	N/A	No	No		
5434-002	1 O/O	Thrombectomy System, Catheter AngioJet Ultra	MEDRAD - Bayer HealthCare Radiology & Interventional (105650-TAB)	Project	Draft (New)	No	No	No	M	15.00	120		Type B (NEMA 5-15)		
				No	2-Movable, Elect	No	No	No	150	23.00	10	N/A	No		
				No	Unassigned	No	No	No	188	52.50	60	No	No		
7986-001	1 O/O	Transilluminator, Venous 05-0100VT (Venous) Height is with lighting system on top of stand.	Electro-Surgical Instrument Co (ESI) (05-0100VT)	Project	Draft (New)	No	No	No	M	30.00	110		Type B (NEMA 5-15)		
				No	2-Movable, Elect	No	No	No	25	30.00	N/A	N/A	No		
				No	Unassigned	No	No	No	32	49.00	60	No	No		
5257-024	1 O/O	Ultrasound, Imaging, Multipurpose, Portable M-Turbo 1.4 Electrical requirements are for battery charging.	FUJIFILM SonoSite, Inc () FUJIFILM SonoSite, Inc ()	Project	Draft (New)	No	No	No	C	11.00	110		Type B (NEMA 5-15)		
				No	2-Movable, Elect	No	No	No	9	12.00	2	N/A	No		
				No	Unassigned	No	No	No	15	3.25	60	No	No		
5740-019	1 O/O	Ultrasound, Imaging, Vascular Access CORE Mobile	Volcano Corp. (400-0100.01 (COREmb120)) Volcano Corp. (400-0100.01 (COREmb120))	Project	Draft (New)	No	No	No	M	22.00	120		Type B (NEMA 5-15)		
				No	2-Movable, Elect	No	No	No	220	33.00	8.3	N/A	Yes		
				No	Unassigned	No	No	No	300	62.00	60	No	No		
7128-010	1 O/O	Warmer, Fluid, Mobile H-1200 Fast Flow w/ Air Detector Clamp Air detector/clamp plugs into mains outlet on fluid warmer.	Smiths Medical - Level 1 Inc. () Smiths Medical - Level 1 Inc. ()	Project	Draft (New)	No	No	No	M	20.00	115		Type B (NEMA 5-15)		
				No	2-Movable, Elect	No	No	No	63	20.00	12	N/A	No		
				No	Unassigned	No	No	No	63	67.00	60	No	No		
4658-011	1 O/O	Warmer, Fluid/ Blood, Portable Gaymar Medi-Temp III	Stryker Medical (FW600)	Project	Draft (New)	No	No	No	C	4.75	120		Type B (NEMA 5-15)		
				Yes	2-Movable, Elect	No	No	No	6	8.75	9	N/A	No		
				No	Unassigned	No	No	No	8	11.25	60	No	No		

Genesis Planning

Memorial Hermann North East Hybrid OR Expansion

Specification Room by Room Detail

Department: Surgery

Atta ID	Qty F/I	Description Model Spec Remarks / Item Notes	Manufacturer Vendor	Funding Source		Item Status Arch Code Custom Code	Water			Vacuum		(lbs) Mounting Max Weight Ship Weight	(in) Width Depth Height	Electrical/Mechanical		
				Arch Sig. Spat. Sig.			Cold Hot Treated	Drain Steam Gas	Dental Medical Vent	Volts Amp Hz	Plug Type BTU/hr Ded. Circ.			Data Cnct. Emrg. Pwr.		
4657-012	1 O/O	Warmer, Patient, Hypothermia Bair Hugger 750	Arizant / 3M Healthcare () Arizant / 3M Healthcare ()	Lease	Draft (New)	No	No	No	C	10.00	110	Type B (NEMA 5-15)				
				Yes	2-Movable, Elect	No	No	No	17	13.50	11.7	N/A	No			
				No	Unassigned	No	No	No	22	12.50	60	No	No			
4687-021	1 O/O	Waste Can, Bio-Hazardous 6145 Red (18 gal)	Rubbermaid Commercial Products () MMS Distributer ()	Project	Draft (New)	No	No	No	F	19.75	N/A	N/A				
				No	3-Movable, Non-Elect	No	No	No	12	16.25	N/A	N/A	No			
				No	Unassigned	No	No	No	13	26.50	N/A	No	No			
5623-013	1 O/O	Waste Disposal, Surgical Fluid Collection Neptune 2 Ultra Rover Height is with IV pole down, Height with IV pole up is 102 inches. Empties via Docking Station (specified separately)	Stryker Instruments () Stryker Instruments ()	Lease	Draft (New)	No	No	No	M	19.00	120	Type B (NEMA 5-15)				
				No	2-Movable, Elect	No	No	No	353	23.00	12	N/A	No			
				No	Unassigned	No	No	No	400	70.00	60	No	No			
4753-015	1 O/V	X-Ray Unit, Interventional, Angio / Cardiac (Single Plane) Allura Xper FD20 Cardiac (Ceiling)	Philips Healthcare - Imaging Systems () Philips Healthcare - Imaging Systems ()	Project	Draft (New)	No	No	No	S	*	480	Hardwire				
				Yes	1-Fixed	No	No	No	*	*	125	N/A	Yes			
				No	Unassigned	No	No	No	*	*	60	No	No			
System consists of multiple components. Refer to Vendor drawings for all site planning information. With Flexmove and 2 Flexvision monitors.																
Room: Operating Room, General (OR5) Room#: Room Sign: Area/Phase: Room Qty: 1																
Comments:																
3370-047	1 O/V	Anesthesia Machine, General Aisys Carestation	GE Healthcare - Anesthesia Delivery (1011-9000-000) GE Healthcare - Anesthesia Delivery (1011-9000-000)	Project	Draft (New)	No	No	No	M	29.00	115	Type B (NEMA 5-15)				
				Yes	2-Movable, Elect	No	No	Yes	370	35.00	15	N/A	Yes			
				No	Unassigned	No	Yes	No	450	59.25	60	No	No			
15 Amp circuit required. Actual amp draw for anesthesia unit only is unavailable (per GE). Unit has on-board electrical outlets for peripherals. Changed model to Aisys. Aestiva no longer available. Changed price from 62,000 to 78,000. kn. 4-29-14																
C-127295	1 C/C	Board, OR Counts Counts Board	WorldWide Medical Solutions () WorldWide Medical Solutions ()	Project	Draft (New)	No	No	No			N/A	N/A				
				No	2-Movable, Elect	No	No	No			N/A	N/A	No			
				No	Unassigned	No	No	No			N/A	No	No			
C-176402	1 O/V	Boom, Anesthesia EC2FVBM36	Skytron () Skytron ()	Project	Draft (New)	No	No	No	CE	N/A	N/A	N/A				
				Yes	2-Movable, Elect	No	No	No	N/A	N/A	N/A	N/A	No			
				No	Unassigned	No	No	No	N/A	N/A	N/A	No	No			
C-187168	1 O/O	Bracket, Monitor CUSTOM Supplied by Surgiquip for Sony monitor to mount to Equipment carrier.	GCX Corporation () SURGIQUIP ()	Project	Draft (New)	No	No	No			N/A	N/A				
				Yes	2-Movable, Elect	No	No	No			N/A	N/A	No			
				No	Unassigned	No	No	No			N/A	No	No			

Genesis Planning

Memorial Hermann North East Hybrid OR Expansion

Specification Room by Room Detail

Department: Surgery

Atta ID	Qty F/I	Description Model Spec Remarks / Item Notes	Manufacturer Vendor	Funding Source		Water			Vacuum (lbs)		(in) Width Depth Height	Electrical/Mechanical		
				Arch Sig. Spat. Sig.	Item Status Arch Code Custom Code	Cold Hot Treated	Drain Steam Gas	Dental Medical Vent	Mounting Max Weight Ship Weight	Volts Amp Hz		Plug Type BTU/hr Ded. Circ.	Data Cnct. Emrg. Pwr.	
3451-001	2 O/O	Bucket, Kick P-1020-SS	Pedigo Products () Pedigo Products ()	Project	Draft (New)	No	No	No	M	13.50	N/A		N/A	
				No	3-Movable, Non-Elect	No	No	No	8	13.50	N/A	N/A	No	
				No	Unassigned	No	No	No	8	14.00	N/A	No	No	
5594-012	1 O/C	Cabinet, OR Console, Supply 48" OR Console (w/sloping top & sub-base) Height includes 8" sloping top and 12" sub-base (optional). Quote Received	Getinge USA - Infection Control () Getinge USA - Infection Control ()	Project	Draft (New)	No	No	No	F	48.00	N/A		N/A	
				Yes	1-Fixed	No	No	No	380	18.00	N/A	N/A	No	
				No	Unassigned	No	No	No	470	80.00	N/A	No	No	
6824-002	1 O/O	Cart, Anesthesia, Controlled Access Anesthesia System Dimensions specified reflect unit fully configured with return bin and top accessory rack. Requires network connection if wireless option not selected. Consult Vendor for project-specific configuration and specification.	CareFusion - Pyxis () CareFusion - Pyxis ()	Lease	Draft (New)	No	No	No	M	47.50	120	Type B (NEMA 5-15)		
				Yes	2-Movable, Elect	No	No	No	400	28.00	3	355	Yes	
				No	Unassigned	No	No	No	450	61.00	60	No	No	
5842-003	1 O/O	Cart, Equipment, General 358 (2 drawers)	LogiQuip LLC () LogiQuip LLC ()	Project	Draft (New)	No	No	No	F	18.75	N/A		N/A	
				No	3-Movable, Non-Elect	No	No	No	44	16.75	N/A	N/A	No	
				No	Unassigned	No	No	No	44	35.00	N/A	No	No	
6364-046	1 O/C	Dispenser, Glove, Triple Box GDSL-T	Sunbelt Plastic () MMS Distributer ()	Project	Draft (New)	No	No	No	W	10.25	N/A		N/A	
				Yes	1-Fixed	No	No	No	2	4.00	N/A	N/A	No	
				No	Unassigned	No	No	No	3	16.00	N/A	No	No	
6457-020	1 O/V	Disposal, Sharps, Floor Cart BioSystems C-17/DWSH-17/LH-17	Stericycle () Stericycle ()	Lease	Draft (New)	No	No	No	M	14.50	N/A		N/A	
				No	3-Movable, Non-Elect	No	No	No	15	22.50	N/A	N/A	No	
				No	Unassigned	No	No	No	18	29.75	N/A	No	No	
4824-060	1 O/V	Electrosurgical Unit, Bipolar ForceTriad	Covidien - Valleylab Div () Covidien - Valleylab Div ()	Project	Draft (New)	No	No	No	C	18.00	120	Type B (NEMA 5-15)		
				Yes	2-Movable, Elect	No	No	No	30	20.00	7	N/A	No	
				Yes	Unassigned	No	No	No	30	10.00	60	No	No	
3836-009	3 O/O	Hamper, Linen P-120-L	Pedigo Products () Pedigo Products ()	Project	Draft (New)	No	No	No	M	19.50	N/A		N/A	
				No	3-Movable, Non-Elect	No	No	No	24	18.00	N/A	N/A	No	
				No	Unassigned	No	No	No	24	32.00	N/A	No	No	
C-190309	1 O/V	Light, Surgical, Dual, Ceiling, w/Monitor Arm & Equip. Boom ECT2F/2AF/AUR55TV	Skytron () Skytron ()	Project	Draft (New)	No	No	No	CE		0		N/A	
				Yes	2-Movable, Elect	No	No	No		N/A	N/A	No		
				No	Unassigned	No	No	No		N/A	No	No		
4076-005	1 O/V	Monitor, Physiologic, Bedside Dinamap MPS Select	GE Healthcare - Monitoring Systems () GE Healthcare - Monitoring Systems ()	Project	Draft (New)	No	No	No	C	16.00	120	Type B (NEMA 5-15)		
				Yes	2-Movable, Elect	No	No	No	40	12.50	N/A	N/A	No	
				No	Unassigned	No	No	No	N/A	15.00	60	No	No	

Genesis Planning

Memorial Hermann North East Hybrid OR Expansion

Specification Room by Room Detail

Department: Surgery

Atta ID	Qty F/I	Description Model Spec Remarks / Item Notes	Manufacturer Vendor	Funding Source		Water			Vacuum		(lbs) Mounting Max Weight Ship Weight	(in) Width Depth Height	Electrical/Mechanical		
				Arch Sig.	Item Status Arch Code Custom Code	Cold Hot Treated	Drain Steam Gas	Dental Medical Vent	Volts Amp Hz	Plug Type BTU/hr Ded. Circ.			Data Cnct. Emrg. Pwr.		
4072-004	1 O/O	Monitor, Physiologic, EEG, Intraoperative BIS Vista Monitor	Covidien RMS (Respiratory & Monitoring) () Covidien - Respiratory and Monitoring ()	Project	Draft (New)	No	No	No	C	7.50	120	Type B (NEMA 5-15)			
				No	2-Movable, Elect	No	No	No	5	5.00	<1	N/A	No		
				No	Unassigned	No	No	No	8	8.00	60	No	Yes		
4248-085	1 O/O	Regulator, Suction, Intermittent/Continuous 8702-1251-908 Push-To-Set Analog	Ohio Medical () MMS Distributer ()	Project	Draft (New)	No	No	No	W	3.00	N/A		N/A		
				No	3-Movable, Non- Elect	No	No	Yes	2	5.00	N/A	N/A	No		
				No	Unassigned	No	No	No	2	6.75	N/A	No	No		
4254-006	1 O/O	Roller, Patient Transfer 920890 (Medium Roller/Rack Kit) Dimensions reflect transfer device hanging on device rack.	Alimed, Inc. (920890) MMS Distributer (920890)	Project	Draft (New)	No	No	No	W	14.00	N/A		N/A		
				No	3-Movable, Non- Elect	No	No	No	10	1.75	N/A	N/A	No		
				No	Unassigned	No	No	No	10	30.00	N/A	No	No		
5795-004	1 O/O	Stand, Basin, Single P-1078-SS	Pedigo Products () Pedigo Products ()	Project	Draft (New)	No	No	No	M	14.00	N/A		N/A		
				No	3-Movable, Non- Elect	No	No	No	15	14.00	N/A	N/A	No		
				No	Unassigned	No	No	No	15	34.00	N/A	No	No		
4363-011	2 O/O	Stand, IV, Multi-Pump iSTAND P-1080-6 (w/Handle & Foley Hook) Height adjustable from 69 to 124 inches. 120V/60Hz 6-outlet strip optional	Pedigo Products () Pedigo Products ()	Project	Draft (New)	No	No	No	M	23.00	120	Type B (NEMA 5-15)			
				No	2-Movable, Elect	No	No	No	47	23.00	N/A	N/A	No		
				No	Unassigned	No	No	No	47	124.00	60	No	No		
4411-017	1 O/O	Stool, Anesthetist P-6000 w/Procedure Rest	Pedigo Products (P-6000/P-6000PR)	Project	Draft (New)	No	No	No	M	21.00	N/A		N/A		
				No	3-Movable, Non- Elect	No	No	No	110	21.00	N/A	N/A	No		
				No	Unassigned	No	No	No	110	42.00	N/A	No	No		
4414-016	2 O/O	Stool, Exam, Cushion-Seat P-526-GS	Pedigo Products () Pedigo Products ()	Project	Draft (New)	No	No	No	M	23.50	N/A		N/A		
				No	3-Movable, Non- Elect	No	No	No	16	23.50	N/A	N/A	No		
				No	Unassigned	No	No	No	16	23.75	N/A	No	No		
5002-001	2 O/O	Stool, Step, Stackable P-1015 Dimensions reflect width and depth with handles. Width and depth without handles: 19"W x 14"D	Pedigo Products () Pedigo Products ()	Project	Draft (New)	No	No	No	F	21.00	N/A		N/A		
				No	3-Movable, Non- Elect	No	No	No	15	16.00	N/A	N/A	No		
				No	Unassigned	No	No	No	15	6.00	N/A	No	No		
5912-002	2 O/O	Table, Instrument, 30-36 inch SG-87-SS (33x18)	Pedigo Products () Pedigo Products ()	Project	Draft (New)	No	No	No	M	33.00	N/A		N/A		
				No	3-Movable, Non- Elect	No	No	No	33	18.00	N/A	N/A	No		
				No	Unassigned	No	No	No	33	34.00	N/A	No	No		

Genesis Planning

Memorial Hermann North East Hybrid OR Expansion

Specification Room by Room Detail

Department: Surgery

Atta ID	Qty F/I	Description Model Spec Remarks / Item Notes	Manufacturer Vendor	Funding Source Arch Sig. Spat. Sig.	Item Status Arch Code Custom Code	Water			Vacuum		(lbs) Mounting Max Weight Ship Weight	(in) Width Depth Height	Electrical/Mechanical		
						Cold Hot Treated	Drain Steam Gas	Dental Medical Vent	Volts Amp Hz	Plug Type BTU/hr Ded. Circ.			Data Cnct. Emrg. Pwr.		
6964-005	1	Table, Surgical, Bariatric	Skytron (6702)	Project	Draft (New)	No	No	No	F	22.00	120	Type B (NEMA 5-15)			
	O/O	6702 HERCULES	Skytron (6702)	Yes	2-Movable, Elect	No	No	No	900	79.00	3.7	N/A No			
		Includes setup.		Yes	Unassigned	No	No	No	1000	41.00	60	No No			
4657-012	1	Warmer, Patient, Hypothermia	Arizant / 3M Healthcare ()	Lease	Draft (New)	No	No	No	C	10.00	110	Type B (NEMA 5-15)			
	O/O	Bair Hugger 750	Arizant / 3M Healthcare ()	Yes	2-Movable, Elect	No	No	No	17	13.50	11.7	N/A No			
				No	Unassigned	No	No	No	22	12.50	60	No No			
4688-016	1	Waste Can, Open Top	Rubbermaid Commercial Products	Project	Draft (New)	No	No	No	F	11.00	N/A	N/A			
	O/O	3541 Slim Jim w/Handles (Gray)	(FG354100LGRAY)	No	3-Movable, Non-Elect	No	No	No	8	23.25	N/A	N/A No			
			MMS Distributer (3541)	No	Unassigned	No	No	No	9	25.00	N/A	No No			
5623-013	1	Waste Disposal, Surgical Fluid	Stryker Instruments ()	Project	Draft (New)	No	No	No	M	19.00	120	Type B (NEMA 5-15)			
	O/O	Collection	Stryker Instruments ()	No	2-Movable, Elect	No	No	No	353	23.00	12	N/A No			
		Neptune 2 Ultra Rover		No	Unassigned	No	No	No	400	70.00	60	No No			
Height is with IV pole down, Height with IV pole up is 102 inches. Empties via Docking Station (specified separately)															
Room: Scrub Sink Room#: Room Sign: Area/Phase: Unassigned Room Qty: 1															
Comments: Scrub sink to be provided by the contractor.															
4688-016	1	Waste Can, Open Top	Rubbermaid Commercial Products	Project	Draft (New)	No	No	No	F	11.00	N/A	N/A			
	O/O	3541 Slim Jim w/Handles (Gray)	(FG354100LGRAY)	No	3-Movable, Non-Elect	No	No	No	8	23.25	N/A	N/A No			
			MMS Distributer (3541)	No	Unassigned	No	No	No	9	25.00	N/A	No No			